



0; dk k; ;  
 211004  
**Purchase Office**  
**Motilal Nehru National Institute of Technology Allahabad, Prayagraj-211004**  
 Tele : 91-532-227+1057, 1030; Email : purchase@mnnit.ac.in



**INVITATION FOR SUPPLY AND INSTALLATION, TESTING, COMMISSIONING OF  
HYDRAULIC SERVO CONTROL UNIVERSAL TESTING MACHINE CAPACITY 1000 kN**

[Open Tender (for goods costing above ₹25 lakhs);

Two Bid System (Simultaneous receipt of Technical & Financial Bids with Technical Bid opening in the first instance and  
Financial Bid opening in the second stage only for the technically responsive bidders)]

| <b>BID SUMMARY</b>   |  |              |  |
|--|--|--------------|--|
|  | Date   | Time         | Venue  |
| Start date & time & Bid submission   | 20.00.2020   | at 1800 Hrs. |  |
| Pre-bid Meeting  | 27.11.2020   | at 1500 Hrs. | Room No. 203, Purchase Office, MNNIT Allahabad, Prayagraj-211004 |
| Closing date & time of Bid submission  | 21.12.2020   | at 1530 Hrs. |  |
| Technical Bid Opening  | 22.12.2020   | at 1530 Hrs. | Room No. 205, Purchase Office, MNNIT Allahabad, Prayagraj-211004 |
| Price Bid Opening  | To be informed later   |              |  |
| Quotations to be submitted   | Online at CPPP viz. <a href="http://www.eprocure.gov.in/eprocure/app">www.eprocure.gov.in/eprocure/app</a> |              |  |
| Earnest Money Deposit (EMD): ₹84,000/- (Eighty Four Thousand only), to be paid through NEFT/RTGS. The Bank details of MNNIT Allahabad is as under:<br><b>Name of the Bank Account:</b> Director, MNNIT Allahabad<br><b>Account No.:</b> 77660100015948<br><b>Name of Bank:</b> Bank of Baroda<br><b>IFSC Code:</b> BARB0VJMNRE (Fifth letter from left is 0 'zero')<br><b>Note:</b> Please upload the transaction receipt. |  |              |  |

Motilal Nehru National Institute of Technology Allahabad invites Online Bids (Two bid system) for the following goods/equipments. The complete Bid Document can be viewed/download from CPPP viz. [www.eprocure.gov.in/eprocure/app](http://www.eprocure.gov.in/eprocure/app) and Institute website [www.mnnit.ac.in](http://www.mnnit.ac.in).

| Sl. | Brief Description of the Items/Equipments   | Qty.   | Place of Delivery          | Installation requirement |
|-----|---|--------|----------------------------|--------------------------|
| 1.  | <b>Hydraulic Servo Control Universal Testing Machine Capacity 1000 kN</b><br>1. Universal testing machine of capacity 1000 kN, designed with PC control electro-hydraulic servo valve loading system & main engine controller separately fixed. It has features like operation amenity, high accuracy, and steady loading. It can perform tests such as tensile, compression, bending, cutting etc. for the metal as well as nonmetallic composite materials. | 01 No. | MNNIT Allahabad, Prayagraj | Yes                      |

2. Technical Specifications:

|                                       |                       |
|---------------------------------------|-----------------------|
| Max. testing force                    | 1000KN                |
| Measuring range                       | 4%~100%F.S            |
| Relative error to force value (%)     | ±1                    |
| Measuring isodromic ratio control     | 0.1-100%F.S./min      |
| Deformation isodromic ratio control   | 0.1-100% F.S./min     |
| Displacement control                  | 0.5-50mm/min          |
| Speed precision                       | ±5% set point         |
| Max. tensile test space               | 700mm                 |
| Flat sample holding thickness         | 0-40mm                |
| Diameter of round specimen            | 20-60mm               |
| Upper and lower platen size (mm)      | 205*205               |
| Bending test pressed rolling distance | 30~600mm              |
| Max. platen rise speed                | ≥70mm/min             |
| Max. moving beam speed                | ≥300mm/min            |
| Motor power                           | ≈2.5KW 380V or Better |
| Noise                                 | ≤75dB or Better       |

3. Electro-Hydraulic measurement Control Cabinet Details:

- Internal micro-computer control panel, communicate with PC by RS232 series port.
- Adopts famous brand import integrated component,
- Adopts integrated high precision meter amplifier.
- High accuracy analog-to-digital conversion.

4. Software Requirement:

- Live indication of load, position, and strain rates, simultaneously during test and peak indication of load, position and strain after test.
- Live X-Y graph of user-selected curve: stress vs. strain; load vs. position; stress vs. position; load vs. strain; load vs. time; position vs. time etc.
- Automatic saving of transducer information to file for later recall, re-graph and recalculation.
- Automatic saving of results to file for statistical computation and export to spreadsheet and database programs: Access and Excel.
- Print single test reports including X-Y graph on a single page
- Select from a pre-defined analysis including: peak load, peak stress, tensile strength, elongation break etc.
- Specimen geometries of flat, round, cylinder, square, tube, pipe.
- Select the international units, such as "N", "kgf", "lb", "mm", "in" and so on.
- Calibration: adjusting hardware of load cell by software, make it meet testing standard.

5. Comply with ISO, BS, ASTM and EN or other standards as ASTM E4, EN10002-2, ISO7500, ISO6892 and BS4449:1998.

|  |   |  |  |  |
|--|---|--|--|--|
|  | <p>6. Grip/Fixture: Versatile removable jaws can be changed rapidly to meet variable wider applications in tension test.</p> <p>7. Laptop and coloured printer with latest configuration.</p> <p>8. Security Protection: Position limiter; emergency switch; the control system can protect automatically for overload, over-voltage, over temperature.</p> |  |  |  |
|--|---|--|--|--|

Bids must be submitted online at CPPP viz. [www.eprocure.gov.in/eprocure/app](http://www.eprocure.gov.in/eprocure/app) and shall be opened in the presence of bidder's authorized representatives, who choose to attend and shall sign a register evidencing their attendance.

The Bid NOT submitted in the prescribed format or Incomplete in any respect shall be outrightly rejected.

No physical Bids shall be submitted in the Institute.

**Assistant Registrar (Admin-III)**  
**MNNIT Allahabad**

## INSTRUCTIONS FOR ONLINE BID SUBMISSION

The bidders are required to submit soft copies of their bids electronically on the CPP Portal, using valid Digital Signature Certificates. The instructions given below are meant to assist the bidders in registering on the CPP Portal, prepare their bids in accordance with the requirements and submitting their bids online on the CPP Portal.

More information useful for submitting online bids on the CPP Portal may be obtained at: <https://eprocure.gov.in/eprocure/app>.

### REGISTRATION:

- 1) Bidders are required to enroll on the e-Procurement module of the Central Public Procurement Portal (URL: <https://eprocure.gov.in/eprocure/app>) by clicking on the link "Online bidder Enrollment" on the CPP Portal which is free of charge.
- 2) As part of the enrolment process, the bidders will be required to choose a unique username and assign a password for their accounts.
- 3) Bidders are advised to register their valid email address and mobile numbers as part of the registration process. These would be used for any communication from the CPP Portal.
- 4) Upon enrolment, the bidders will be required to register their valid Digital Signature Certificate (Class III Certificates with signing key usage) issued by any Certifying Authority recognized by CCA India (e.g. Sify / nCode / eMudhra etc.), with their profile.
- 5) Only one valid DSC should be registered by a bidder. Please note that the bidders are responsible to ensure that they do not lend their DSC's to others which may lead to misuse.
- 6) Bidder then logs in to the site through the secured log-in by entering their user ID/password and the password of the DSC/e-Token.

### SEARCHING FOR TENDER DOCUMENTS:

- 1) There are various search options built in the CPP Portal, to facilitate bidders to search active tenders by several parameters. These parameters could include Tender ID, Organization Name, Location, Date, Value, etc. There is also an option of advanced search for tenders, wherein the bidders may combine a number of search parameters such as Organization Name, Form of Contract, Location, Date, Other keywords etc. to search for a tender published on the CPP Portal.
- 2) Once the bidders have selected the tenders they are interested in, they may download the required documents / tender schedules. These tenders can be moved to the respective 'My Tenders' folder. This would enable the CPP Portal to intimate the bidders through SMS / e- mail in case there is any corrigendum issued to the tender document.
- 3) The bidder should make a note of the unique Tender ID assigned to each tender, in case they want to obtain any clarification / help from the Helpdesk.

### PREPARATION OF BIDS:

- 1) Bidder should take into account any corrigendum published on the tender document before submitting their bids.
- 2) Please go through the tender advertisement and the tender document carefully to understand the documents required to be submitted as part of the bid. Please note the number of covers in which the bid documents have to be submitted, the number of documents - including the names and content of each of the document that need to be submitted. Any deviations from these may lead to rejection of the bid.
- 3) Bidder, in advance, should get ready the bid documents to be submitted as indicated in the tender document / schedule and generally, they can be in PDF / XLS / RAR / DWF/JPG formats. Bid documents may be scanned with 100 dpi with black and white option which helps in reducing size of the scanned document.
- 4) To avoid the time and effort required in uploading the same set of standard documents which are required to be submitted as a part of every bid, a provision of uploading such standard documents (e.g. PAN card copy, annual reports, auditor certificates etc.) has been provided to the bidders. Bidders can use "My

Space" or "Other Important Documents" area available to them to upload such documents. These documents may be directly submitted from the "My Space" area while submitting a bid, and need not be uploaded again and again. This will lead to a reduction in the time required for bid submission process.

**Note:** *My Documents space is only a repository given to the Bidders to ease the uploading process. If Bidder has uploaded his Documents in My Documents space, this does not automatically ensure these Documents being part of Technical Bid.*

### **SUBMISSION OF BIDS:**

- 1) Bidder should log into the site well in advance for bid submission so that they can upload the bid in time i.e. on or before the bid submission time. Bidder will be responsible for any delay due to other issues.
- 2) The bidder has to digitally sign and upload the required bid documents one by one as indicated in the tender document.
- 3) Bidder has to select the payment option as "offline" to pay the tender fee / EMD as applicable and enter details of the instrument.
- 4) Bidder should prepare the EMD as per the instructions specified in the tender document.

The original should be posted/couriered/given in person to the concerned official, latest by the last date of bid submission or as specified in the tender documents. The details of the DD/any other accepted instrument, physically sent, should tally with the details available in the scanned copy and the data entered during bid submission time. Otherwise the uploaded bid will be rejected.

- 5) Bidders are requested to note that they should necessarily submit their financial bids in the format provided and no other format is acceptable. If the price bid has been given as a standard BoQ format with the tender document, then the same is to be downloaded and to be filled by all the bidders. Bidders are required to download the BoQ file, open it and complete the white coloured (unprotected) cells with their respective financial quotes and other details (such as name of the bidder). No other cells should be changed. Once the details have been completed, the bidder should save it and submit it online, without changing the filename. If the BoQ file is found to be modified by the bidder, the bid will be rejected.
- 6) The server time (which is displayed on the bidders' dashboard) will be considered as the standard time for referencing the deadlines for submission of the bids by the bidders, opening of bids etc. The bidders should follow this time during bid submission.
- 7) All the documents being submitted by the bidders would be encrypted using PKI encryption techniques to ensure the secrecy of the data. The data entered cannot be viewed by unauthorized persons until the time of bid opening. The confidentiality of the bids is maintained using the secured Socket Layer 128 bit encryption technology. Data storage encryption of sensitive fields is done. Any bid document that is uploaded to the server is subjected to symmetric encryption using a system generated symmetric key. Further this key is subjected to asymmetric encryption using buyers/bid opener's public keys. Overall, the uploaded tender documents become readable only after the tender opening by the authorized bid openers.
- 8) The uploaded tender documents become readable only after the tender opening by the authorized bid openers.
- 9) Upon the successful and timely submission of bids (i.e. after Clicking "Freeze Bid Submission" in the portal), the portal will give a successful bid submission message & a bid summary will be displayed with the bid no. and the date & time of submission of the bid with all other relevant details.
- 10) The bid summary has to be printed and kept as an acknowledgement of the submission of the bid. This acknowledgement may be used as an entry pass for any bid opening meetings.

**ASSISTANCE TO BIDDERS:**

- 1) Any queries relating to the tender document and the terms and conditions contained therein should be addressed to the Tender Inviting Authority for a tender or the relevant contact person indicated in the tender.
- 2) Any queries relating to the process of online bid submission or queries relating to CPP Portal in general may be directed to the 24x7 CPP Portal Helpdesk.

## GENERAL TERMS & CONDITION (GCC)

### 1. Bid Price:

- (a) The contract shall be for the full quantity as described above. Corrections, if any, shall be made by crossing out, initialing, dating and rewriting.
- (b) All duties, taxes and other levies payable by the contractor under the contract shall be included in the total price.
- (c) The rates quoted by the bidder shall be fixed for the duration of the contract and shall not be subject to adjustment on any account.
- (d) The prices should be quoted in Indian Rupees only for indigenous items. For imported items/Equipments, the prices should be quoted in foreign currency.
- (e) Each bidder shall submit only one quotation.

### 2. Validity of Quotation:

Quotation shall remain valid for a period not less than **180 days** after the deadline date specified for submission.

### 3. Formats and Signing of Bid:

- Each & every pages of the submitted bid must carry the page numbers.
- The bidders are not allowed to make additional and alteration in the tender document, any additions and alteration in the tender document, any additions and alterations shall be at bidder's risk. Conditional/modified tender are liable to be rejected.
- The tender/quotations NOT submitted in the prescribed format or Incomplete in any respect will be outrightly rejected.
- If the bid submitted without compliance sheet details of equipments the bid will be outrightly rejected.
- The bid shall be typed or written in indelible ink and shall be signed by the Bidder/Tenderer or a person or persons duly authorized to bind the Bidder/Tenderer to the Contract. All pages of the bid, except for unamended printed literature, shall be initialed by the person or persons signing the bid.
- Any interlineations, erasure or overwriting shall be valid only if they are initiated by the persons or persons signing the bid.
- The Bidder shall furnish information as described in the Form of Bid on commissions or gratuities, if any, paid or to be paid to agents relating to this Bid, and to contract execution if the Bidder is awarded the contract.

### 4. Eligibility Criteria:

- (i) The Bidder must be a legally valid entity i.e. a Proprietary/Partnership Firm/Limited Company/Society Legally constituted or registered under the relevant act
- (ii) Bidder must have an annual average turnover of **₹135.00 lakhs** during the past three financial years i.e. 2017-18 and 2018-19 and 2019-20.
- (iii) Must have supplied the same equipments with same set of accessories to reputed Indian organizations especially NITS/IITs/IIITs/Central Universities/IISERs/CSIR/Labs etc. during the last three financial years i.e. 2017-18 and 2018-19 and 2019-20.
- (iv) Duly filled in format 'A' to 'F'.
- (v) Must be conforming to the terms and conditions, and specifications of the indented item (s).
- (vi) Must have submitted the EMD required unless exempted.

**NOTE:** The bidders/tenderers who finally emerge as technically acceptable RESPONSIVE BIDDERS will be considered for opening of PRICE BID and will be intimated for the same through Institute's website <http://mnnit.ac.in/tenders.html> and [www.eprocure.gov.in/eprocure/app](http://www.eprocure.gov.in/eprocure/app).

5. **Award of Contract:**

The Purchaser will award the contract to the bidder whose quotation has been determined to be substantially responsive and who has offered the lowest evaluated quotation price.

- Notwithstanding the above, the Purchaser reserves the right to accept or reject any quotations and to cancel the bidding process and reject all quotations at any time prior to the award of contract.
- The bidder whose bid is accepted will be notified for the award of contract by the Purchaser prior to expiration of the quotation validity period. The terms of the accepted offer shall be incorporated in the purchase order if falls under Govt. purchase rules, i.e., GFR-2017

6. **Delivery:**

02 months from the date of receipt of Purchase Order.

7. **Liquidated Damages:**

While granting extension of the delivery period, where the delivery of stores or any installment thereof is accepted after expiry of the original delivery period, the CA may recover from the contractor, as agreed, the LD a sum equivalent to 0.5 (Half) per cent of the prices of any portion of stores delivered late, for each week or part thereof of delay. The total damages shall not exceed 10 (Ten) percent of the value of delayed goods. The LD cannot exceed the amount stipulated in the contract.

8. **Force Majeure:**

A Force Majeure (FM) means extraordinary events or circumstance beyond human control such as an event described as an act of God (like a natural calamity) or events such as a war, strike, riots, crimes (but not including negligence or wrong-doing, predictable/seasonal rain and any other events specifically excluded in the clause). An FM clause in the contract frees both parties from contractual liability or obligation when prevented by such events from fulfilling their obligations under the contract. An FM clause does not excuse a party's non-performance entirely, but only suspends it for the duration of the FM. The firm has to give notice of FM as soon as it occurs and it cannot be claimed ex-post facto. There may be a FM situation affecting the purchase organisation only. In such a situation, the purchase organisation is to communicate with the supplier along similar lines as above for further necessary action. If the performance in whole or in part or any obligation under this contract is prevented or delayed by any reason of FM for a period exceeding 90 (Ninety) days, either party may at its option terminate the contract without any financial repercussion on either side.

9. **Payment condition:**

Payment condition: 100% Payment shall be released only after delivery, installation & satisfactory acceptance of goods. PBG as noted in para 5 of Instruction/Guidelines to Bidder (ITB) shall from part of the condition for release of payment. **Advance payment is not permitted.**

i) Normal Commercial warranty/guarantee shall be applicable to the supplied goods.

ii) **Payment will be cleared through e-payment mode (*Format of e-payment form given at Annexure-F*)**



10. **Criteria for Technical Assessment for would be as follows:**

1. Preference will be given to ISO 9001:2000 certified Manufacturer/suppliers, who can ensure the manufacturing of the machine as per the required testing standards IS/BS/BIS/ASTM standards and within the specified tolerance limits.
2. Manufactures should have NABL accredited testing and calibration facility from minimum last five years.
3. Bidders should provide list of plant and machineries and inspection tools along with quotation.
4. Manufacturer should have in-hours R & D center duly recognized by Department of Scientific & Industrial Research, Ministry of Science and Technology. Bidder has to provide the certificate along with the tender.
5. Manufacturer will ensure the tractability (source of calibration) of the calibration unit to NCCBM, NPL etc. Bidders should quote make of the quoted equipment along with authorized letter from manufacturer.
6. Before placing the orders, supplier should offer us factory visit to ensure the capability of supplier.
7. Bidders should provide copies of original Memorandum and Articles of Association, defining the constitution or legal status, place or registration and place of business of the company.
8. Bidders should provide details of service centre and information on service support facilities that would be provided after the warranty period. Manufacturer should have trained and qualified customer support staff with ample experience in the required field.
9. Offer should be submitted with Original Technical Literature by bidder.
10. Warranty period should be two years from the date of installation.

### INSTRUCTIONS/GUIDELINES TO BIDDERS (ITB)

1. Bidders who had supplied the Government departments, PSUs/NITs/IITs/IIITs and reputed Institutions and Organizations should only apply.
2. The bidder should enclose the client list with contact address along with Phone/Fax for reference, for the quoted item.
3. Please quote the prices of F.O.R. Motilal Nehru National Institute of Technology Allahabad, Prayagraj.

**[NOTE: Custom/Excise Duty Exemption Certificate and GST Exemption Certificate will be provided by MNNIT Allahabad after receiving your request letter for the same.**

**(i) MNNIT Allahabad avails Custom Duty Exemption in terms of Government Notification N. 51/96-Customs dated 23 July, 1996, and Central Excise Duty Exemption in terms of Government Notification N. 10/97-Central Excise dated 01 March, 1997, as amended from time to time.**

**(ii) MNNIT Allahabad avails GST Concession under Notification No. 47/2017 Central Tax (Rate) dated 14 November, 2017, as amended from time to time.**

4. The Bidders are requested to submit their Quotation online alongwith proof of submission of EMD. Quotation without EMD will be summarily rejected. EMD's of all Bidders will be returned back after the purchase order /contract agreement is issued to successful bidder.
5. Successful Bidders must have to submit the performance security @10% of the purchase order or contract value and it will be retained up to the warranty period.
6. The bidder should enclose relevant documents wherever necessary to substantiate his eligibility.
7. The Institute is eligible for and can provide necessary documentation for Custom/Excise and Octroi exemption on demand, if permitted under rule.
8. Motilal Nehru National Institute of Technology Allahabad, Prayagraj, reserves the right to reject, accept any or whole or part of any of the tender without assigning any reason thereof and no claim will be heard. In case of dispute, the decision of The Director, MNNIT, Allahabad will be final and binding.
9. If the successful bidder is not able to supply the goods items within the delivery date specified in the purchase order, the purchase order will be automatically treated as cancelled after expiry of delivery date period. The vendor will have to submit an undertaking in this regard that this condition is acceptable to him. In case of extra ordinary circumstances the vendors must send a request for extension of validity of purchase orders, with proper justification prior to the expiry of validity date for consideration.
10. In case of sophisticated equipments, an on-site training needs to be provided with no extra cost.
11. The quotation must be accompanied by leaflets/operation manuals/books etc.
12. Wherever applicable the vendors must quote the branded /desired items.

**CHECK LIST  
(ON THE LETTER HEAD OF THE BIDDER)**

The Bid will be the compilation of following documents, along with required supporting documents. No document in support of minimum eligibility criteria will be accepted / entertained after opening of tender.

| <b>Sl. No</b> | <b>Documents</b>   |
|---------------|--|
| 1.            | Cover letter by bidder (On the Letter Head of the Bidder)  |
| 2.            | Earnest Money Deposit (EMD), if required   |
| 3.            | 02 (Two) nos. of recently executed relevant Purchase Orders for the same/similar product to different central/state government organizations/ PSU/Autonomous bodies owned by Central/State Government. |
| 4.            | <b>Annexure-A</b> : Bid Proposal Sheet (On the Letter Head of the Bidder)  |
| 5.            | <b>Annexure-B</b> : Compliance of bidder with reference to equipments (On the Letter Head of the Bidder)   |
| 6.            | <b>Annexure-C</b> : Proforma for Price Reasonability Certificate (On The Letter Head Of The Bidder)  |
| 7.            | <b>Annexure-D</b> : Manufacturers' Authorization Certificate   |
| 8.            | <b>Annexure-E</b> : Affidavit regarding Blacklisting/Non-Blacklisting of firm  |
| 9.            | <b>Annexure-F</b> : Proforma for Direct Payment/Transfer To Bank Account By MNNIT Allahabad  |

**BID PROPOSAL SHEET**  
**(ON THE LETTER HEAD OF THE BIDDER)**  
**[To be uploaded]**

To  
 Assistant Registrar (Admin-III)  
 MNNIT Allahabad, Prayagraj

**Subject:** Supply, Installation, Testing and Commissioning of Hydraulic Servo Control Universal Testing Machine Capacity 1000 kN.

Dear Sir,

We, the undersigned having read and examined in detail the specifications as specified in this document in respect of Supply, Installation, Testing and Commissioning of Hydraulic Servo Control Universal Testing Machine Capacity 1000 kN at MNNIT Allahabad, Prayagraj do hereby propose to supply the required products and services.

|               |      |  |         |
|---------------|------|--|---------|
| EMD submitted |      | YES / NO<br>(Please strike off whatever is not applicable) |         |
| Amount        | Mode | Date of payment  | UTR No. |
|               |      |  |         |

- (i) **ADDITIONAL PURCHASE/WORK ORDER:** We understand that the MNNIT Allahabad, Prayagraj, in case of the requirements may also place repeat purchase order/work order on the company. In such cases, we shall accept and execute all the purchase/work order placed on us by MNNIT Allahabad, Prayagraj.
- (ii) **BID PRICING:** We further declare that the prices stated in our proposal are in accordance with your Terms & Conditions in the bidding document. We further understand that the quantities as specified in this Tender may increase or decrease at the time of Award of Purchase Order as per the requirements of MNNIT Allahabad, Prayagraj.
- (iii) **QUALIFYING DATA:** We confirm that we satisfy the qualifying criteria and have attached the requisite documents as documentary proofs. In case you require any further information/documentary proof in this regard during evaluation of our bid, we agree to furnish the same in time to your satisfaction.
- (iv) **CONTRACT PERFORMANCE SECURITY:** We hereby declare that in case the contract is awarded to us, we shall submit the performance Guarantee Bond in the form of Bank Guarantee for the amount mentioned at **clause No. 5 in ITB** of the total order value.
- (v) **PAYMENT TERMS:** We hereby declare that in case the contract is awarded to us, we agree with payment terms specified in the tender documents.
- (vi) **CERTIFICATE AND DECLARATION:**
- I/We certify that no addition/modification/alteration has been made in the Original Tender Document. If at any stage addition /modification /alteration is noticed in the Original Document, I/We will abide by the terms and conditions contained in the original tender document, failing which MNNIT Allahabad reserves the right to reject the tender and/or cancel the contract
  - It has been certified that all information provided in tender form is true and correct to the best of my knowledge and belief. We hereby declare that our proposal is made in good faith, without collusion or fraud. No forged /tampered document(s) are produced with tender form for gaining unlawful advantage. We understand that MNNIT Allahabad is authorized to make enquiry to establish the facts claimed and obtained confidential reports from clients.
  - In case it is established that any information provided by us is false / misleading or in the circumstances where it is found that we have made any wrong claims. Further MNNIT Allahabad is also authorized to blacklist our firm/company/agency and debar us in participating in any tender/bid in future.
  - I/We assure the Institute that neither I/We, nor any of my/our workers, will do any act which is improper/illegal during the execution in case the tender is awarded to us.
  - I/We assure the Institute that I/We will NOT be outsourcing any work specified in the tender document,

to any other firm.

- f) Neither I/We, nor anybody on my/our behalf will indulge in any corrupt activities/practices in my/our dealing with the Institute.
- g) Our Firm/Company/Agency is not been blacklisted or banned by any Govt. Department, PSU, University, Autonomous Institute or any other Govt. Organization.
- h) I/We certify that, I have understood all the terms & conditions (GCC and ITB etc.), as indicated in enquiry of the tender document, and hereby accept all the same completely.
- i) I/We, further certify that I/We, possess all the statutory/non-statutory registrations, permissions, approvals, etc., from the Competent Authority for providing the requisite services,
- j) We understand that you are not bound to accept the lowest or any bid you may receive.
- k) I/We hereby declare that this tender on acceptance communicated by you shall constitute a valid and binding contract between us.
- l) I/We certify that the submitted quotation is duly paginated and contains from page no. 1 to .....

Date:

**Technical Compliance of Goods  
(ON THE LETTER HEAD OF THE BIDDER)**

[To be uploaded]

| Sl.                                   | Brief Description of the Items/Equipments   | Qty.               | Compliance<br>(Yes/No) | Quoted<br>Model |            |                                   |    |                                   |                  |                                     |                   |                      |              |                 |               |                         |       |                               |        |                            |         |                                  |         |                                       |          |                        |           |                        |            |             |                       |       |                 |        |  |  |
|---------------------------------------|---|--------------------|------------------------|-----------------|------------|-----------------------------------|----|-----------------------------------|------------------|-------------------------------------|-------------------|----------------------|--------------|-----------------|---------------|-------------------------|-------|-------------------------------|--------|----------------------------|---------|----------------------------------|---------|---------------------------------------|----------|------------------------|-----------|------------------------|------------|-------------|-----------------------|-------|-----------------|--------|--|--|
| 1.                                    | <p><b>Hydraulic Servo Control Universal Testing Machine Capacity 1000 kN</b></p> <p>1. Universal testing machine of capacity 1000 kN, designed with PC control electro-hydraulic servo valve loading system &amp; main engine controller separately fixed. It has features like operation amenity, high accuracy, and steady loading. It can perform tests such as tensile, compression, bending, cutting etc. for the metal as well as nonmetallic composite materials.</p> <p>2. Technical Specifications:</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tbody> <tr> <td>Max. testing force</td> <td>1000KN</td> </tr> <tr> <td>Measuring range</td> <td>4%~100%F.S</td> </tr> <tr> <td>Relative error to force value (%)</td> <td>±1</td> </tr> <tr> <td>Measuring isodromic ratio control</td> <td>0.1-100%F.S./min</td> </tr> <tr> <td>Deformation isodromic ratio control</td> <td>0.1-100% F.S./min</td> </tr> <tr> <td>Displacement control</td> <td>0.5-50mm/min</td> </tr> <tr> <td>Speed precision</td> <td>±5% set point</td> </tr> <tr> <td>Max. tensile test space</td> <td>700mm</td> </tr> <tr> <td>Flat sample holding thickness</td> <td>0-40mm</td> </tr> <tr> <td>Diameter of round specimen</td> <td>20-60mm</td> </tr> <tr> <td>Upper and lower platen size (mm)</td> <td>205*205</td> </tr> <tr> <td>Bending test pressed rolling distance</td> <td>30~600mm</td> </tr> <tr> <td>Max. platen rise speed</td> <td>≥70mm/min</td> </tr> <tr> <td>Max. moving beam speed</td> <td>≥300mm/min</td> </tr> <tr> <td>Motor power</td> <td>≈2.5KW 380V or Better</td> </tr> <tr> <td>Noise</td> <td>≤75dB or Better</td> </tr> </tbody> </table> <p>3. Electro-Hydraulic measurement Control Cabinet Details:</p> <ul style="list-style-type: none"> <li>• Internal micro-computer control panel, communicate with PC by RS232 series port.</li> <li>• Adopts famous brand import integrated component,</li> <li>• Adopts integrated high precision meter amplifier.</li> <li>• High accuracy analog-to-digital conversion.</li> </ul> <p>4. Software Requirement:</p> <ul style="list-style-type: none"> <li>• Live indication of load, position, and strain rates, simultaneously during test and peak indication of load, position and strain after test.</li> <li>• Live X-Y graph of user-selected curve: stress vs. strain; load vs. position; stress vs. position; load vs. strain; load vs. time; position vs. time etc.</li> <li>• Automatic saving of transducer information to file for later recall, re-graph and recalculation.</li> <li>• Automatic saving of results to file for statistical computation</li> </ul> | Max. testing force | 1000KN                 | Measuring range | 4%~100%F.S | Relative error to force value (%) | ±1 | Measuring isodromic ratio control | 0.1-100%F.S./min | Deformation isodromic ratio control | 0.1-100% F.S./min | Displacement control | 0.5-50mm/min | Speed precision | ±5% set point | Max. tensile test space | 700mm | Flat sample holding thickness | 0-40mm | Diameter of round specimen | 20-60mm | Upper and lower platen size (mm) | 205*205 | Bending test pressed rolling distance | 30~600mm | Max. platen rise speed | ≥70mm/min | Max. moving beam speed | ≥300mm/min | Motor power | ≈2.5KW 380V or Better | Noise | ≤75dB or Better | 01 No. |  |  |
| Max. testing force                    | 1000KN  |                    |                        |                 |            |                                   |    |                                   |                  |                                     |                   |                      |              |                 |               |                         |       |                               |        |                            |         |                                  |         |                                       |          |                        |           |                        |            |             |                       |       |                 |        |  |  |
| Measuring range                       | 4%~100%F.S  |                    |                        |                 |            |                                   |    |                                   |                  |                                     |                   |                      |              |                 |               |                         |       |                               |        |                            |         |                                  |         |                                       |          |                        |           |                        |            |             |                       |       |                 |        |  |  |
| Relative error to force value (%)     | ±1  |                    |                        |                 |            |                                   |    |                                   |                  |                                     |                   |                      |              |                 |               |                         |       |                               |        |                            |         |                                  |         |                                       |          |                        |           |                        |            |             |                       |       |                 |        |  |  |
| Measuring isodromic ratio control     | 0.1-100%F.S./min  |                    |                        |                 |            |                                   |    |                                   |                  |                                     |                   |                      |              |                 |               |                         |       |                               |        |                            |         |                                  |         |                                       |          |                        |           |                        |            |             |                       |       |                 |        |  |  |
| Deformation isodromic ratio control   | 0.1-100% F.S./min   |                    |                        |                 |            |                                   |    |                                   |                  |                                     |                   |                      |              |                 |               |                         |       |                               |        |                            |         |                                  |         |                                       |          |                        |           |                        |            |             |                       |       |                 |        |  |  |
| Displacement control                  | 0.5-50mm/min  |                    |                        |                 |            |                                   |    |                                   |                  |                                     |                   |                      |              |                 |               |                         |       |                               |        |                            |         |                                  |         |                                       |          |                        |           |                        |            |             |                       |       |                 |        |  |  |
| Speed precision                       | ±5% set point   |                    |                        |                 |            |                                   |    |                                   |                  |                                     |                   |                      |              |                 |               |                         |       |                               |        |                            |         |                                  |         |                                       |          |                        |           |                        |            |             |                       |       |                 |        |  |  |
| Max. tensile test space               | 700mm   |                    |                        |                 |            |                                   |    |                                   |                  |                                     |                   |                      |              |                 |               |                         |       |                               |        |                            |         |                                  |         |                                       |          |                        |           |                        |            |             |                       |       |                 |        |  |  |
| Flat sample holding thickness         | 0-40mm  |                    |                        |                 |            |                                   |    |                                   |                  |                                     |                   |                      |              |                 |               |                         |       |                               |        |                            |         |                                  |         |                                       |          |                        |           |                        |            |             |                       |       |                 |        |  |  |
| Diameter of round specimen            | 20-60mm   |                    |                        |                 |            |                                   |    |                                   |                  |                                     |                   |                      |              |                 |               |                         |       |                               |        |                            |         |                                  |         |                                       |          |                        |           |                        |            |             |                       |       |                 |        |  |  |
| Upper and lower platen size (mm)      | 205*205   |                    |                        |                 |            |                                   |    |                                   |                  |                                     |                   |                      |              |                 |               |                         |       |                               |        |                            |         |                                  |         |                                       |          |                        |           |                        |            |             |                       |       |                 |        |  |  |
| Bending test pressed rolling distance | 30~600mm  |                    |                        |                 |            |                                   |    |                                   |                  |                                     |                   |                      |              |                 |               |                         |       |                               |        |                            |         |                                  |         |                                       |          |                        |           |                        |            |             |                       |       |                 |        |  |  |
| Max. platen rise speed                | ≥70mm/min   |                    |                        |                 |            |                                   |    |                                   |                  |                                     |                   |                      |              |                 |               |                         |       |                               |        |                            |         |                                  |         |                                       |          |                        |           |                        |            |             |                       |       |                 |        |  |  |
| Max. moving beam speed                | ≥300mm/min  |                    |                        |                 |            |                                   |    |                                   |                  |                                     |                   |                      |              |                 |               |                         |       |                               |        |                            |         |                                  |         |                                       |          |                        |           |                        |            |             |                       |       |                 |        |  |  |
| Motor power                           | ≈2.5KW 380V or Better   |                    |                        |                 |            |                                   |    |                                   |                  |                                     |                   |                      |              |                 |               |                         |       |                               |        |                            |         |                                  |         |                                       |          |                        |           |                        |            |             |                       |       |                 |        |  |  |
| Noise                                 | ≤75dB or Better   |                    |                        |                 |            |                                   |    |                                   |                  |                                     |                   |                      |              |                 |               |                         |       |                               |        |                            |         |                                  |         |                                       |          |                        |           |                        |            |             |                       |       |                 |        |  |  |

|  |  |  |  |  |
|--|--|--|--|--|
|  | <p>and export to spreadsheet and database programs: Access and Excel.</p> <ul style="list-style-type: none"> <li>• Print single test reports including X-Y graph on a single page</li> <li>• Select from a pre-defined analysis including: peak load, peak stress, tensile strength, elongation break etc.</li> <li>• Specimen geometries of flat, round, cylinder, square, tube, pipe.</li> <li>• Select the international units, such as "N", "kgf", "lb", "mm", "in" and so on.</li> <li>• Calibration: adjusting hardware of load cell by software, make it meet testing standard.</li> </ul> <p>5. Comply with ISO, BS, ASTM and EN or other standards as ASTM E4, EN10002-2, ISO7500, ISO6892 and BS4449:1998.</p> <p>6. Grip/Fixture: Versatile removable jaws can be changed rapidly to meet variable wider applications in tension test.</p> <p>7. Laptop and coloured printer with latest configuration.</p> <p>8. Security Protection: Position limiter; emergency switch; the control system can protect automatically for overload, over-voltage, over temperature.</p> |  |  |  |
|--|--|--|--|--|

**PROFORMA FOR PRICE REASONABILITY CERTIFICATE**  
**(Company's/Firm Letterhead)**  
**PRICE REASONABILITY CERTIFICATE**  
**[To be uploaded]**

It is certified that the rates quoted against Enquiry No. .... dated .....for the supply of items vide our Quotation No. ....dated ..... for (Currency/Value)..... addressed to MNNIT Allahabad are exclusively for supply to ACADEMIC and RESEARCH Institutions and are not more than as charged to other Government /PSU's for same supplies made in recent past.

We would like to certify that the quoted price are the minimum and neither we have quoted the same item on lesser rates, than those being offered to MNNIT Allahabad, to any other customer for same supplies made in recent past, and nor we will do so till the validity of offer or execution of purchase order, whichever is later.

If these rates are approved by the Director, MNNIT Allahabad and if at any stage the quoted rates are found higher than the rates applicable to supply to Government, then in such condition MNNIT Allahabad will have the right to cancel the approved rates/Purchase order and to take legal action against the tenderer.

Yours faithfully,  
(Signature with date)

Name:

Designation:

Seal:

Duly authorized to sign tender/enquiry for and on behalf of

---

---



MANUFACTURERS' AUTHORIZATION CERTIFICATE

No.....

dated.....

To,

Dear Sir:

We.....who are established and reputable manufacturers of (<Name and descriptions of goods offered>) having factories at.....(<Address of factory>) do hereby authorize M/s.....(<Name/address of Agent>) to submit a bid, and sign the contract with you for the goods manufactured by us against the above Enquiry.

We hereby extend our full guarantee and warranty for the goods and services offered for supply by the above firm against this Enquiry.

The undersigned is authorized to issue this 'Manufacturers' Authorization Certificate' on behalf of the company/firm <M/s.....name of company/firm .....>.

Yours faithfully,

Signature of the authorized signatory

Name & Designation of the authorized signatory

Contact No. & E-Mail of the authorized signatory

Company/Firm's Seal

**Note:** This letter of authority should be signed by a person competent and having the power of attorney to legally bind the manufacturer. It should be included by the Bidder in its bid.

**AFFIDAVIT REGARDING BLACKLISTING/ NON-BLACKLISTING FROM TAKING PART IN GOVT.****TENDER BY GOVT**

**(To be executed on Rs.10/- Stamp paper & attested by Public Notary/Executive Magistrate by the bidder)**

I/We Proprietor/Partner(s)/Director(s) of M/S. .... hereby declare that the firm/company namely M/S.....has not been blacklisted or debarred in the past by MNNIT Allahabad or any other Government organization from taking part in Government tenders.

Or

I/We Proprietor/Partner(s)/Director(s) of M/S. .... hereby declare that the firm/company namely M/S.....was blacklisted or debarred by MNNIT Allahabad, or any other Government Department from taking part in Government tenders for a period of ----- years w.e.f.-----.

The period is over on-----and now the firm/company is entitled to take part in Government tenders. In case the above information found false I/We are fully aware that the tender/ contract will be rejected/cancelled by MNNIT Allahabad, and EMD/SD shall be forfeited. In addition to the above MNNIT Allahabad, will not be responsible to pay the bills for any completed/ partially completed work.

Signature .....

Name.....

Capacity in which assigned: .....

Name & address of the firm: .....

Date:

Signature of Bidder with seal

**ANNEXURE-F****PROFORMA FOR DIRECT PAYMENT/TRANSFER TO BANK ACCOUNT BY MNNIT ALLAHABAD**

**[To be uploaded]**

| Sl. | Particulars  | Information |
|-----|--|-------------|
| 1.  | Firm [Beneficiary] Name  |             |
| 2.  | Complete Bank Account No. of the Firm [beneficiary].<br>[in case of change in bank account vendor should write to Accounts Office] |             |
| 3.  | Bank Name  |             |
| 4.  | Branch Address   |             |
| 5.  | IFSC Code No.  |             |
| 6.  | Permanent Account Number   |             |
| 7.  | Mobile No. [for SMS]   |             |
| 8.  | Email ID [for information]   |             |

We undertake that all information provided above is correct and MNNIT Allahabad will not be responsible in case of any error on the part of firm.

[Seal and Signature of the firm]